

General Dehydroepiandrosterone (DHEA) ELISA Kit | DHEA elisa kit

General Dehydroepiandrosterone (DHEA) ELISA Kit

CATALOG #
AAA122462

REACTIVITY
General


PRODUCT SYNONYMS

Dehydroepiandrosterone (DHEA); General Dehydroepiandrosterone (DHEA) ELISA Kit; DHEA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Assay Type	Quantitative Sandwich
Samples	Serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids
Detection Range	0.05-20nmol/ml
Sensitivity	0.023nmol/ml
Intra-assay Precision	Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision. Intra-Assay: CV<8%
Inter-assay Precision	Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision. CV(%) = SD/mean x 100. Inter-Assay: CV<10%
Preparation and Storage	Store at 2 to 8 degree C for 6 months.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for DHEA elisa kit

Principle of the Assay: This kit is an Enzyme-Linked Immunosorbent Assay (ELISA). The plate has been pre-coated with General DHEA antibody. DHEA present in the sample is added and binds to antibodies coated on the wells. And then biotinylated General DHEA Antibody is added and binds to DHEA in the sample. Then Streptavidin-HRP is added and binds to the Biotinylated DHEA antibody. After incubation unbound Streptavidin-HRP is washed away during a washing step. Substrate solution is then added and color develops in proportion to the amount of General DHEA. The reaction is terminated by addition of acidic stop solution and absorbance is measured at 450 nm.

Intended Uses: This sandwich kit is for the accurate quantitative detection of General Dehydroepiandrosterone (also known as DHEA) in serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/general-elisa-kits/dehydroepiandrosterone-dhea/122462>