

## Human Cluster of Differentiation 155 ELISA Kit

Human Cluster of Differentiation 155 ELISA Kit

CATALOG #  
**AAA122983**

REACTIVITY  
**Human**


### PRODUCT SYNONYMS

Cluster of Differentiation 155; Human Cluster of Differentiation 155 ELISA Kit; Cluster of Differentiation 155 elisa kit

### PRODUCT SPECIFICATIONS

|                         |                                                                                                                                                                                                  |
|-------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Reactivity              | Human                                                                                                                                                                                            |
| Samples                 | Serum, Plasma, Cell Culture Supernates, Cell Lysates, Tissue Homogenates                                                                                                                         |
| Assay Type              | Quantitative Sandwich                                                                                                                                                                            |
| Detection Range         | 0.05ng/ml-15ng/ml                                                                                                                                                                                |
| Sensitivity             | 0.017ng/ml                                                                                                                                                                                       |
| Intra-assay Precision   | Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision. Intra-Assay: CV<8%                              |
| Inter-assay Precision   | Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision. CV(%) = SD/mean x 100. Inter-Assay: CV<10% |
| Preparation and Storage | Store at 2 to 8 degree C for 6 months.                                                                                                                                                           |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for Cluster of Differentiation 155 elisa kit

**Intended Uses:** This sandwich kit is for the accurate quantitative detection of human Cluster of Differentiation 133 (also known as CD133) in serum, plasma, cell culture supernates, cell lysates, tissue homogenates.

**Principle of the Assay:** This kit is an Enzyme-Linked Immunosorbent Assay (ELISA). The plate has been pre-coated with human CD133 antibody. CD133 present in the sample is added and binds to antibodies coated on the wells. And then biotinylated human CD133 Antibody is added and binds to CD133 in the sample. Then Streptavidin-HRP is added and binds to the Biotinylated CD133 antibody. After incubation unbound Streptavidin-HRP is washed away during a washing step. Substrate solution is then added and color develops in proportion to the amount of human CD133. The reaction is terminated by addition of acidic stop solution and absorbance is measured at 450 nm.

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | [contact@aaabiotech.com](mailto:contact@aaabiotech.com)

Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/cluster-of-differentiation-155/122983>