

Mouse Left/Right Determination Factor 1 (LEFTY1) CLIA Kit | LEFTY1 clia kit

Mouse Left/Right Determination Factor 1 (LEFTY1) CLIA Kit

CATALOG #
AAA129966

REACTIVITY
Mouse


PRODUCT SYNONYMS

Left/Right Determination Factor 1 (LEFTY1); Mouse Left/Right Determination Factor 1 (LEFTY1) CLIA Kit; LEFTY-B; LEFTYB; LEFT-B; LEFTB; Left Right Determination Factor B; LEFTY1 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of Left/Right Determination Factor 1 (LEFTY1). No significant cross-reactivity or interference between Left/Right Determination Factor 1 (LEFTY1) and analogues was observed.
Sequence Length	366
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue homogenates and other biological fluids
Detection Range	13.7-10,000pg/mL
Sensitivity	5.6pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Left/Right Determination Factor 1 (LEFTY1) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Left/Right Determination Factor 1 (LEFTY1) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for LEFTY1 clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Left/Right Determination Factor 1 (LEFTY1). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Left/Right Determination Factor 1 (LEFTY1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Left/Right Determination Factor 1 (LEFTY1) level in the sample or standard.;
Product Categories/Family for LEFTY1 clia kit	Neuroscience

NCBI AND UNIPROT INFORMATION

NCBI GI #	10337603
NCBI GeneID	10637
NCBI Accession #	NP_066277.1
NCBI GenBank Nucleotide #	NM_020997.3
UniProt Accession #	O75610; Q53H67; Q5TE94; B2R7U0
Molecular Weight	40,880 Da
NCBI Official Full Name	left-right determination factor 1 preproprotein
NCBI Official Synonym Full Names	left-right determination factor 1
NCBI Official Symbol	LEFTY1
NCBI Official Synonym Symbols	LEFTB; LEFTYB
NCBI Protein Information	left-right determination factor 1
UniProt Protein Name	Left-right determination factor 1
UniProt Gene Name	LEFTY1
UniProt Synonym Gene Names	LEFTB; LEFTYB
UniProt Entry Name	LFTY1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.