

Human Angiotensin II (AngII) CLIA Kit | AngII clia kit

Human Angiotensin II (AngII) CLIA Kit

CATALOG #
AAA130165

REACTIVITY
Human


PRODUCT SYNONYMS

Angiotensin II (AngII); Human Angiotensin II (AngII) CLIA Kit; Ang-II; Angiotensin-2; AngII clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Angiotensin II (AngII). No significant cross-reactivity or interference between Angiotensin II (AngII) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, Plasma, Tissue homogenates, Cell lysates, Cell culture supernates and other biological fluids
Detection Range	11.72-3,000pg/mL
Sensitivity	3.61pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Angiotensin II (AngII) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Angiotensin II (AngII) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for AngII clia kit	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Angiotensin II (AngII). A competitive inhibition reaction is launched between biotin labeled Angiotensin II (AngII) and unlabeled Angiotensin II (AngII) (Standards or samples) with the pre-coated antibody specific to Angiotensin II (AngII). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Angiotensin II (AngII) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Angiotensin II (AngII) level in the sample or standard.
Product Categories/Family for AngII clia kit	Endocrinology; Cardiovascular biology

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-clia-kits/angiotensin-ii-angii/130165>