

Human Interleukin 20 (IL20) CLIA Kit | IL20 clia kit

Human Interleukin 20 (IL20) CLIA Kit

CATALOG #
AAA130447

REACTIVITY
Human


PRODUCT SYNONYMS

Interleukin 20 (IL20); Human Interleukin 20 (IL20) CLIA Kit; IL10D; IL10-D; ZCYTO10; Cytokine Zcyto10; IL20 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Interleukin 20 (IL20). No significant cross-reactivity or interference between Interleukin 20 (IL20) and analogues was observed.
Sequence Length	176
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and other biological fluids
Detection Range	1.37-1,000pg/mL
Sensitivity	0.61pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Interleukin 20 (IL20) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Interleukin 20 (IL20) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IL20 clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Interleukin 20 (IL20). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Interleukin 20 (IL20). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Interleukin 20 (IL20) level in the sample or standard.;
Product Categories/Family for IL20 clia kit	Cytokine ; Infection immunity ; Dermatology

NCBI AND UNIPROT INFORMATION

NCBI GI #	109658898
NCBI GeneID	50604
UniProt Accession #	Q9NYY1 ; Q17RB3 ; Q2THG6 ; Q96QZ6
Molecular Weight	17,144 Da
NCBI Official Full Name	Interleukin 20
NCBI Official Synonym Full Names	interleukin 20
NCBI Official Symbol	IL20
NCBI Official Synonym Symbols	IL-20; IL10D; ZCYTO10
NCBI Protein Information	interleukin-20
UniProt Protein Name	Interleukin-20
UniProt Gene Name	IL20
UniProt Synonym Gene Names	ZCYTO10; IL-20
UniProt Entry Name	IL20_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.