

## Human Keratin 33A (KRT33A) ELISA Kit | KRT33A elisa kit

Human Keratin 33A (KRT33A) ELISA Kit

CATALOG #  
**AAA136413**

REACTIVITY  
**Human**


### PRODUCT SYNONYMS

Keratin 33A (KRT33A); Human Keratin 33A (KRT33A) ELISA Kit; KRTHA3A; Ha-3I; Krt1-3; K33A; HHA3-I; HKA3A; Keratin; Hair; Acidic 3A; Hard Keratin Type I 3I; type I cuticular Ha3-I; KRT33A elisa kit

### PRODUCT SPECIFICATIONS

|                                                   |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |
|---------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Reactivity                                        | Human                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
| Specificity                                       | This assay has high sensitivity and excellent specificity for detection of Keratin 33A (KRT33A). No significant cross-reactivity or interference between Keratin 33A (KRT33A) and analogues was observed.                                                                                                                                                                                                                                                                                  |
| Assay Type                                        | Double-antibody Sandwich                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| Samples                                           | Tissue homogenates, Cell lysates, Cell culture supernates and other biological fluids                                                                                                                                                                                                                                                                                                                                                                                                      |
| Detection Range                                   | 0.156-10ng/mL                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| Sensitivity                                       | 0.061ng/mL                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |
| Application                                       | Enzyme-linked immunosorbent assay for Antigen Detection.                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| Intra-assay Precision (Precision within an assay) | 3 samples with low, middle and high level Keratin 33A (KRT33A) were tested 20 times on one plate, respectively.                                                                                                                                                                                                                                                                                                                                                                            |
| Inter-assay Precision (Precision between assays)  | 3 samples with low, middle and high level Keratin 33A (KRT33A) were tested on 3 different plates, 8 replicates in each plate.                                                                                                                                                                                                                                                                                                                                                              |
| CV(%) =                                           | SD/meanX100                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
| Intra-Assay                                       | CV<10%                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |
| Inter-Assay                                       | CV<12%                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |
| Preparation and Storage                           | The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.<br>To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end. |

### IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

|                                                  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
|--------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Related Product Information for KRT33A elisa kit | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Keratin 33A (KRT33A). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Keratin 33A (KRT33A). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Keratin 33A (KRT33A), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Keratin 33A (KRT33A) in the samples is then determined by comparing the O.D. of the samples to the standard curve. |
| Product Categories/Family for KRT33A elisa kit   | Tumor immunity                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | [contact@aaabiotech.com](mailto:contact@aaabiotech.com)  
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/keratin-33a-krt33a/136413>