

Mouse Heparin Cofactor II (HCII) ELISA Kit | HCII elisa kit

Mouse Heparin Cofactor II (HCII) ELISA Kit

CATALOG #
AAA138881

REACTIVITY
Mouse


PRODUCT SYNONYMS

Heparin Cofactor II (HCII); Mouse Heparin Cofactor II (HCII) ELISA Kit; SERPIND1; HC2; HCF2; HC II; HLS2; LS2; Serpin D1; Serpin Peptidase Inhibitor Clade D Member 1; Protease inhibitor leuserpin-2; HCII elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of Heparin Cofactor II (HCII). No significant cross-reactivity or interference between Heparin Cofactor II (HCII) and analogues was observed.
Sequence Length	499
Assay Type	Double-antibody Sandwich
Samples	Plasma
Detection Range	0.156-10ng/mL
Sensitivity	0.069ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Heparin Cofactor II (HCII) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Heparin Cofactor II (HCII) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HCII elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Heparin Cofactor II (HCII). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Heparin Cofactor II (HCII). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Heparin Cofactor II (HCII), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Heparin Cofactor II (HCII) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for HCII elisa kit	Hematology

NCBI AND UNIPROT INFORMATION

NCBI GI #	183910
NCBI GeneID	3053
UniProt Accession #	P05546; Q6IBZ5; B2RAI1; D3DX34
Molecular Weight	57,071 Da
NCBI Official Full Name	heparin cofactor II
NCBI Official Synonym Full Names	serpin peptidase inhibitor, clade D (heparin cofactor), member 1
NCBI Official Symbol	SERPIND1
NCBI Official Synonym Symbols	HC2; LS2; HCF2; HCII; HLS2; THPH10; D22S673
NCBI Protein Information	heparin cofactor 2
UniProt Protein Name	Heparin cofactor 2
UniProt Gene Name	SERPIND1
UniProt Synonym Gene Names	HCF2; HC-II; HLS2
UniProt Entry Name	HEP2_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.