

Human Macrophage Migration Inhibitory Factor (MIF) ELISA Kit | MIF elisa kit

Human Macrophage Migration Inhibitory Factor (MIF) ELISA Kit

CATALOG #
AAA140431

REACTIVITY
Human


PRODUCT SYNONYMS

Macrophage Migration Inhibitory Factor (MIF); Human Macrophage Migration Inhibitory Factor (MIF) ELISA Kit; GIF; GLIF; MMIF; Glycosylation-Inhibiting Factor; L-dopachrome isomerase; L-dopachrome tautomerase; Phenylpyruvate tautomerase; MIF elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Macrophage Migration Inhibitory Factor (MIF). No significant cross-reactivity or interference between Macrophage Migration Inhibitory Factor (MIF) and analogues was observed.
Sequence Length	115
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue homogenates, Cell lysates, Cell culture supernates and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	0.122ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Macrophage Migration Inhibitory Factor (MIF) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Macrophage Migration Inhibitory Factor (MIF) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MIF elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Macrophage Migration Inhibitory Factor (MIF). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Macrophage Migration Inhibitory Factor (MIF). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Macrophage Migration Inhibitory Factor (MIF), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Macrophage Migration Inhibitory Factor (MIF) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for MIF elisa kit	Cytokine; Infection immunity; Rheumatology; Autoimmunity

NCBI AND UNIPROT INFORMATION

NCBI GI #	4505185
NCBI GeneID	4282
NCBI Accession #	NP_002406.1
NCBI GenBank Nucleotide #	NM_002415.1
UniProt Accession #	P14174; Q2V4Y5; Q6FHV0; A5Z1R8; B2R4S3
Molecular Weight	12,476 Da
NCBI Official Full Name	macrophage migration inhibitory factor
NCBI Official Synonym Full Names	macrophage migration inhibitory factor (glycosylation-inhibiting factor)
NCBI Official Symbol	MIF
NCBI Official Synonym Symbols	GIF; GLIF; MMIF
NCBI Protein Information	macrophage migration inhibitory factor
UniProt Protein Name	Macrophage migration inhibitory factor
UniProt Gene Name	MIF
UniProt Synonym Gene Names	GLIF; MMIF; MIF; GIF
UniProt Entry Name	MIF_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.