

Human Crystallin Lambda 1 (CRYI1) ELISA Kit | CRYI1 elisa kit

Human Crystallin Lambda 1 (CRYI1) ELISA Kit

CATALOG #
AAA140536

REACTIVITY
Human


PRODUCT SYNONYMS

Crystallin Lambda 1 (CRYI1); Human Crystallin Lambda 1 (CRYI1) ELISA Kit; GDH; Lambda-CRY; Gul3DH; L-Gulonate 3-Dehydrogenase; Lambda-Crystallin Homolog; CRYI1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Crystallin Lambda 1 (CRYI1). No significant cross-reactivity or interference between Crystallin Lambda 1 (CRYI1) and analogues was observed.
Sequence Length	265
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	0.063ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Crystallin Lambda 1 (CRYI1) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Crystallin Lambda 1 (CRYI1) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CRY11 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Crystallin Lambda 1 (CRY11). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Crystallin Lambda 1 (CRY11). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Crystallin Lambda 1 (CRY11), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Crystallin Lambda 1 (CRY11) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for CRY11 elisa kit	Immune molecule

NCBI AND UNIPROT INFORMATION

NCBI GI #	119628666
NCBI GeneID	51084
Molecular Weight	33,359 Da
NCBI Official Full Name	crystallin, lambda 1, isoform CRA_c
NCBI Official Synonym Full Names	crystallin, lambda 1
NCBI Official Symbol	CRYL1
NCBI Official Synonym Symbols	GDH; HEL30; lambda-CRY
NCBI Protein Information	lambda-crystallin homolog
UniProt Protein Name	Lambda-crystallin homolog
UniProt Gene Name	CRYL1
UniProt Synonym Gene Names	CRY; Gul3DH
UniProt Entry Name	CRYL1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.