

General Vanillylmandelic Acid (VMA) ELISA Kit | VMA elisa kit

General Vanillylmandelic Acid (VMA) ELISA Kit

CATALOG #
AAA148409

REACTIVITY
General


PRODUCT SYNONYMS

Vanillylmandelic Acid (VMA); General Vanillylmandelic Acid (VMA) ELISA Kit; Vanillyl Mandelic Acid; VMA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of VMA. No significant cross-reactivity or interference between VMA and analogues was observed.
Assay Type	Quantitative Competitive
Samples	Urine and biological agents
Detection Range	6.17-500ng/mL
Sensitivity	< 2.94ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level VMA were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level VMA were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for VMA elisa kit	Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to VMA has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled VMA and unlabeled VMA (Standards or samples) with the pre-coated antibody specific to VMA. After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of VMA in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of VMA in the sample. Intended Uses: The kit is a competitive inhibition enzyme immunoassay technique for the in vitro quantitative measurement of VMA in urine and biological agents.
Product Categories/Family for VMA elisa kit	Metabolic pathway; Endocrinology; Cardiovascular biology; Hepatology; Gastroenterology; Hormone metabolism

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/general-elisa-kits/vanillylmandelic-acid-vma/148409>