

Mouse Anti-Mullerian Hormone (AMH) CLIA Kit | AMH clia kit

Mouse Anti-Mullerian Hormone (AMH) CLIA Kit

CATALOG #
AAA149964

REACTIVITY
Mouse


PRODUCT SYNONYMS

Anti-Mullerian Hormone (AMH); Mouse Anti-Mullerian Hormone (AMH) CLIA Kit; MIF; MIH; MIS; Müllerian Inhibiting Factor; Müllerian Inhibiting Hormone; Müllerian Inhibiting Substance; AMH clia kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of Anti-Mullerian Hormone (AMH). No significant cross-reactivity or interference between Anti-Mullerian Hormone (AMH) and analogues was observed.
Sequence Length	603
Assay Type	Competitive Inhibition
Samples	Serum, Plasma, Tissue homogenates, Cell lysates, Cell culture supernates and other biological fluids
Detection Range	39.06-10,000pg/mL
Sensitivity	14.13pg/mL
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Anti-Mullerian Hormone (AMH) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Anti-Mullerian Hormone (AMH) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for AMH clia kit	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Anti-Mullerian Hormone (AMH). A competitive inhibition reaction is launched between biotin labeled Anti-Mullerian Hormone (AMH) and unlabeled Anti-Mullerian Hormone (AMH) (Standards or samples) with the pre-coated antibody specific to Anti-Mullerian Hormone (AMH). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Anti-Mullerian Hormone (AMH) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Anti-Mullerian Hormone (AMH) level in the sample or standard.
Product Categories/Family for AMH clia kit	Endocrinology; Reproductive Science; Hormone Metabolism

NCBI AND UNIPROT INFORMATION

NCBI GI #	557355818
NCBI GeneID	108706872
NCBI Official Full Name	anti-Mullerian hormone
NCBI Official Synonym Full Names	anti-Mullerian hormone S homeolog
NCBI Official Symbol	amh.S
NCBI Official Synonym Symbols	AMH; mif; mis
NCBI Protein Information	muellerian-inhibiting factor-like

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/mouse-clia-kits/anti-mullerian-hormone-amh/149964>