

Human Clara cell protein, CC16 ELISA Kit | CC16 elisa kit

Human Clara cell protein, CC16 ELISA Kit

CATALOG #
AAA15435

REACTIVITY
Human


PRODUCT SYNONYMS

Clara cell protein; CC16; Human Clara cell protein; CC16 ELISA Kit; CC10; CCSP; UGB; SCGB1A1; Uteroglobin (Clara-cell specific 10-kD protein); secretoglobin; family 1A; member 1 (uteroglobin) ; CC16 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of human CC16. No significant cross-reactivity or interference between human CC16 and analogues was observed.
Assay Type	Quantitative Sandwich
Samples	Serum, plasma, tissue homogenates
Detection Range	0.625 ng/ml-40 ng/ml
Sensitivity	< 0.156 ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess.
Inter-assay Precision	Inter-assay Precision (Precision between assays): CV%<10%. Three samples of known concentration were tested in twenty assays to assess.
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CC16 elisa kit

Principle of the Assay: This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for CC16 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any CC16 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for CC16 is added to the wells. After washing, avidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of CC16 bound in the initial step. The color development is stopped and the intensity of the color is measured.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/clara-cell-protein-cc16/15435>