

Human oxidized high density lipoprotein (Ox-HDL) ELISA Kit | Ox-HDL elisa kit

Human oxidized high density lipoprotein, Ox-HDL ELISA Kit

CATALOG #
AAA15824

REACTIVITY
Human


PRODUCT SYNONYMS

oxidized high density lipoprotein (Ox-HDL); Human oxidized high density lipoprotein; Ox-HDL ELISA Kit; Human oxidized high density lipoprotein (Ox-HDL) ELISA Kit; Ox-HDL elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of human Ox-HDL. No significant cross-reactivity or interference between human Ox-HDL and analogues was observed.
Assay Type	Quantitative Sandwich
Samples	Serum, plasma, tissue homogenates, urine
Detection Range	31.25 pg/ml-2000 pg/ml
Sensitivity	< 7.81 pg/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess.
Inter-assay Precision	Inter-assay Precision (Precision between assays): CV%<10%. Three samples of known concentration were tested in twenty assays to assess.
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Ox-HDL elisa kit

Principle of the Assay: This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for Ox-HDL has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any Ox-HDL present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for Ox-HDL is added to the wells. After washing, avidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of Ox-HDL bound in the initial step. The color development is stopped and the intensity of the color is measured.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/oxidized-high-density-lipoprotein-ox-hdl/15824>