

Human Fibrinogen Degradation Product ELISA Kit | FDP elisa kit

Human Fibrinogen Degradation Product ELISA Kit

CATALOG #
AAA16187

REACTIVITY
Human


PRODUCT SYNONYMS

Fibrinogen Degradation Product; Human Fibrinogen Degradation Product ELISA Kit; FDP elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	5.0-100ng/mL
Sensitivity	1.0ng/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for FDP elisa kit	Intended Uses: This FDP ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human FDP. This ELISA kit for research use only! Principle of the Assay: FDP ELISA kit applies the competitive enzyme immunoassay technique utilizing a polyclonal anti-FDP antibody and an FDP-HRP conjugate. The assay sample and buffer are incubated together with FDP-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the FDP concentration since FDP from samples and FDP-HRP conjugate compete for the anti-FDP antibody binding site. Since the number of sites is limited, as more sites are occupied by FDP from the sample, fewer sites are left to bind FDP-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The FDP concentration in each sample is interpolated from this standard curve.
Product Categories/Family for FDP elisa kit	Cardiovascular

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/fdp-human-elisa-kits/fibrinogen-degradation-product/16187>