

Human CRYalphaB ELISA Kit | CRYalphaB elisa kit

Human CRYalphaB (Crystallin Alpha B) ELISA Kit

CATALOG #
AAA163934

REACTIVITY
Human


PRODUCT SYNONYMS

CRYalphaB; Human CRYalphaB (Crystallin Alpha B) ELISA Kit; CRYalphaB elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This kit recognizes Human CRYalphaB in samples. No significant cross-reactivity or interference between Human CRYalphaB and analogues was observed.
Samples	Serum, plasma and other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	0.16-10ng/mL
Sensitivity	0.1ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Human CRYalphaB were tested 20 times on one plate, respectively.
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Human CRYalphaB were tested on 3 different plates, 20 replicates in each plate.
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CRYalphaB elisa kit	Intended Uses: This ELISA kit applies to the in vitro quantitative determination of Human CRYalphaB concentrations in serum, plasma and other biological fluids. Principle of the Assay: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human CRYalphaB. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human CRYalphaB and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human CRYalphaB, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The OD value is proportional to the concentration of Human CRYalphaB. You can calculate the concentration of Human CRYalphaB in the samples by comparing the OD of the samples to the standard curve.
---	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/cryalphab/163934>