

Human beta-TG ELISA Kit | beta-TG elisa kit

Human beta-TG (beta-Thromboglobulin) ELISA Kit

CATALOG #
AAA164991

REACTIVITY
Human


PRODUCT SYNONYMS

beta-TG; Human beta-TG (beta-Thromboglobulin) ELISA Kit; beta-TG elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This kit recognizes Human betaTG/PBP/CXCL7/NAP2 in samples. No significant crossreactivity or interference between Human betaTG/PBP/CXCL7/NAP2 and analogues was observed
Assay Type	Quantitative Sandwich
Samples	Serum, plasma and other biological fluids
Detection Range	31.25-2000 pg/mL
Sensitivity	18.75 pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Human betaTG/PBP/CXCL7/NAP2 were tested 20 times on one plate, respectively.
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Human betaTG/PBP/CXCL7/NAP2 were tested on 3 different plates, 20 replicates in each plate, respectively.
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for beta-TG elisa kit	Principle of the Assay: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human betaTG/PBP/CXCL7/NAP2. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human betaTG/PBP/CXCL7/NAP2 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human betaTG/PBP/CXCL7/NAP2, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzymesubstrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 ± 2 nm. The OD value is proportional to the concentration of Human betaTG/PBP/CXCL7/NAP2. You can calculate the concentration of Human betaTG/PBP/CXCL7/NAP2 in the samples by comparing the OD of the samples to the standard curve. Intended Uses: This ELISA kit applies to the in vitro quantitative determination of Human betaTG/PBP/CXCL7/NAP2 concentrations in serum, plasma and other biological fluids.
---	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/beta-tg/164991>