

Rat JAK2 ELISA Kit | JAK2 elisa kit

Rat JAK2 (Janus Kinase 2) ELISA Kit

CATALOG #
AAA165815

REACTIVITY
Rat


PRODUCT SYNONYMS

JAK2; Rat JAK2 (Janus Kinase 2) ELISA Kit; JAK2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This kit recognizes natural and recombinant Rat JAK2. No significant cross-reactivity or interference between Rat JAK2 and analogues was observed.
Samples	Serum, Plasma, Biological Fluids
Assay Type	Sandwich
Detection Range	0.156-10ng/mL
Sensitivity	Min: 0.094ng/mL; Max: 10ng/mL
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for JAK2 elisa kit

Intended Uses: This ELISA kit applies to the invitro quantitative determination of Rat JAK2 concentrations in serum, plasma and other biological fluids.

Principle of the Assay: This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to JAK2. Standards or samples are added to the appropriate micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for JAK2 and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain JAK2, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The OD value is proportional to the concentration of JAK2. You can calculate the concentration of JAK2 in the samples by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/rat-elisa-kits/jak2/165815>