

Rat Pro-ADM ELISA Kit | Pro-ADM elisa kit

Rat Pro-ADM(Proadrenomedullin)ELISA Kit

CATALOG #
AAA165842

REACTIVITY
Rat


PRODUCT SYNONYMS

Pro-ADM; Rat Pro-ADM(Proadrenomedullin)ELISA Kit; Pro-ADM elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This kit recognizes natural and some recombinant Rat Pro-ADM. No significant crossreactivity or interference between Rat Pro-ADM and analogues was observed.
Samples	Serum, Plasma, Biological Fluids
Assay Type	Sandwich
Detection Range	1.563-100pmol/mL
Sensitivity	Min: 0.938pmol/mL; Max: 100pmol/mL
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Pro-ADM elisa kit

Intended Uses: This ELISA kit can be applied to the in vitro quantitative determination of Rat Pro-ADM concentrations in serum, plasma and other biological fluids.

Principle of the Assay: This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit has been precoated with an antibody specific to Rat Pro-ADM. Standards or samples are added to appropriate micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibodies specific for Rat Pro-ADM and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well successively and incubated. After incubation, free components are washed away. Then the Substrate Reagent is added to each well, only those wells that contain Rat Pro-ADM, biotinylated detection antibody and Avidin- HRP conjugate will appear blue in color. The enzyme-substrate reaction will be terminated by adding Stop Solution and appears yellow in color. The optical density (OD) can be measured with spectrophotometry at a wavelength of 450 nm +/- 2 nm. The OD value is proportional to the concentration of Rat Pro-ADM. The concentration of Rat Pro-ADM in samples can be calculated by comparing the OD of the samples with the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/rat-elisa-kits/pro-adm/165842>