

Human 3Alpha Androstenediol ELISA Kit | 3Alphadiol elisa kit

Human 3Alpha Androstenediol ELISA Kit

CATALOG #
AAA17267

REACTIVITY
Human


PRODUCT SYNONYMS

3Alpha Androstenediol; Human 3Alpha Androstenediol ELISA Kit; 3Alphadiol elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	1.0-25ng/mL
Sensitivity	0.1ng/mL
Intended Uses	This 3alphaAS ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human 3alphaAS. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for 3Alphadiol elisa kit

Principle of the assay: 3alphaAS ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-3alphaAS antibody and an 3alphaAS-HRP conjugate. The assay sample and buffer are incubated together with 3alphaAS-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of color is inversely proportional to the 3alphaAS concentration since 3alphaAS from samples and 3alphaAS-HRP conjugate compete for the anti-3alphaAS antibody binding site. Since the number of sites is limited, as more sites are occupied by 3alphaAS from the sample, fewer sites are left to bind 3alphaAS-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The 3alphaAS concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

