

Human Cardiotrophin 1 (CT1) ELISA Kit | CT1 elisa kit

Human Cardiotrophin 1 (CT1) ELISA Kit

CATALOG #
AAA183997

REACTIVITY
Human


PRODUCT SYNONYMS

Cardiotrophin 1 (CT1); Human Cardiotrophin 1 (CT1) ELISA Kit; CTF1; CT1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Cardiotrophin 1 (CT1). No significant cross-reactivity or interference between Cardiotrophin 1 (CT1) and analogues was observed.
Sequence Length	201
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue Homogenates, Cell Lysates and Other Biological Fluids
Detection Range	15.6-1,000pg/mL
Sensitivity	< 5.9pg/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Cardiotrophin 1 (CT1) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Cardiotrophin 1 (CT1) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CT1 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Cardiotrophin 1 (CT1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Cardiotrophin 1 (CT1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Cardiotrophin 1 (CT1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Cardiotrophin 1 (CT1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	4503121
NCBI GeneID	1489
NCBI Accession #	NP_001321.1
NCBI GenBank Nucleotide #	NM_001330.3
UniProt Accession #	Q16619; Q5U5Y7; A8MVX4
Molecular Weight	21,098 Da
NCBI Official Full Name	cardiotrophin-1 isoform 1
NCBI Official Synonym Full Names	cardiotrophin 1
NCBI Official Symbol	CTF1
NCBI Official Synonym Symbols	CT1; CT-1
NCBI Protein Information	cardiotrophin-1
UniProt Protein Name	Cardiotrophin-1
UniProt Gene Name	CTF1
UniProt Synonym Gene Names	CT-1

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.