

General Dopamine (DA) CLIA Kit | DA clia kit

General Dopamine (DA) CLIA Kit


CATALOG # AAA188435	REACTIVITY General	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	------------------------------	---

PRODUCT SYNONYMS

Dopamine (DA); General Dopamine (DA) CLIA Kit; Dopamine; 2-(3; 4-Dihydroxyphenyl)ethylamine; 3; 4-Dihydroxyphenethylamine; 3-Hydroxytyramine; Intropin; Revivan; Oxytyramine; Dopamine Hydrochloride; DA clia kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of Dopamine (DA). No significant cross-reactivity or interference between Dopamine (DA) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, plasma and other biological fluids
Detection Range	11.72-3,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 4.91pg/mL
Applicable Applications for DA clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Dopamine (DA) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Dopamine (DA) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Dopamine (DA). A competitive inhibition reaction is launched between biotin labeled Dopamine (DA) and unlabeled Dopamine (DA) (Standards or samples) with the pre-coated antibody specific to Dopamine (DA). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Dopamine (DA) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Dopamine (DA) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/general-clia-kits/dopamine-da/188435>