

Human Albumin (ALB) CLIA Kit | ALB clia kit

Human Albumin (ALB) CLIA Kit

CATALOG # AAA188445	REACTIVITY Human	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	----------------------------	---


PRODUCT SYNONYMS

Albumin (ALB); Human Albumin (ALB) CLIA Kit; Albumin; Serum albumin; ALB clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Albumin (ALB). No significant cross-reactivity or interference between Albumin (ALB) and analogues was observed.
Sequence Length	609
Assay Type	Competitive Inhibition
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	1.56-400ug/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.52ug/mL
Applicable Applications for ALB clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Albumin (ALB) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Albumin (ALB) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Albumin (ALB). A competitive inhibition reaction is launched between biotin labeled Albumin (ALB) and unlabeled Albumin (ALB) (Standards or samples) with the pre-coated antibody specific to Albumin (ALB). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Albumin (ALB) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Albumin (ALB) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	4502027
NCBI GeneID	213
NCBI Accession #	NP_000468.1
NCBI GenBank Nucleotide #	NP_000468.1
UniProt Accession #	P02768
NCBI Official Full Name	serum albumin preproprotein
NCBI Official Synonym Full Names	albumin
NCBI Official Symbol	ALB
NCBI Official Synonym Symbols	HSA; PRO0883; PRO0903; PRO1341
NCBI Protein Information	serum albumin
UniProt Protein Name	Serum albumin
UniProt Gene Name	ALB
UniProt Entry Name	ALBU_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.