

Chicken Luteinizing Hormone (LH) ELISA Kit | LH elisa kit

Chicken Luteinizing Hormone (LH) High Sensitive ELISA Kit


CATALOG # AAA188492	REACTIVITY Chicken	APPLICATIONS ELISA
-------------------------------	------------------------------	------------------------------

PRODUCT SYNONYMS

Luteinizing Hormone (LH); Chicken Luteinizing Hormone (LH) High Sensitive ELISA Kit; Luteinizing Hormone; ICSH; Lutropin; Interstitial Cell Stimulating Hormone; LH elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Chicken
Specificity	This assay has high sensitivity and excellent specificity for detection of High Sensitive Luteinizing Hormone (LH). No significant cross-reactivity or interference between High Sensitive Luteinizing Hormone (LH) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, plasma and other biological fluids
Detection Range	123.5-10,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 58.5pg/mL
Applicable Applications for LH elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level High Sensitive Luteinizing Hormone (LH) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level High Sensitive Luteinizing Hormone (LH) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Test Principle	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to High Sensitive Luteinizing Hormone (LH) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled High Sensitive Luteinizing Hormone (LH) and unlabeled High Sensitive Luteinizing Hormone (LH) (Standards or samples) with the pre-coated antibody specific to High Sensitive Luteinizing Hormone (LH). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of High Sensitive Luteinizing Hormone (LH) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of High Sensitive Luteinizing Hormone (LH) in the sample.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/chicken-elisa-kits/luteinizing-hormone-lh/188492>