

Human Hypoxia Inducible Factor 1 Alpha (HIF1a) ELISA Kit | HIF1a elisa kit

Human Hypoxia Inducible Factor 1 Alpha (HIF1a) High Sensitive ELISA Kit

CATALOG # AAA188518	REACTIVITY Human	APPLICATIONS ELISA
-------------------------------	----------------------------	------------------------------


PRODUCT SYNONYMS

Hypoxia Inducible Factor 1 Alpha (HIF1a); Human Hypoxia Inducible Factor 1 Alpha (HIF1a) High Sensitive ELISA Kit; Hypoxia Inducible Factor 1 Alpha; HIF1-A; MOP1; PASD8; Basic Helix-Loop-Helix Transcription Factor; ARNT-interacting protein; Basic-helix-loop-helix-PAS protein MOP1; Class E basic helix-loop-helix protein 78; HIF1a elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of HIF1a. No significant cross-reactivity or interference between HIF1a and analogues was observed.
Sequence Length	826
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	31.2-2,000pg/mL
Sensitivity	< 13.9pg/mL
Applicable Applications for HIF1a elisa kit	ELISA
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level HIF1a were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level HIF1a were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HIF1a elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of HIF1a in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids. Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to HIF1a. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to HIF1a. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain HIF1a, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of HIF1a in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	4504385
NCBI GeneID	3091
NCBI Accession #	NP_001521.1
NCBI GenBank Nucleotide #	NP_001521.1
UniProt Accession #	Q16665
NCBI Official Full Name	hypoxia-inducible factor 1-alpha isoform 1
NCBI Official Synonym Full Names	hypoxia inducible factor 1 subunit alpha
NCBI Official Symbol	HIF1A
NCBI Official Synonym Symbols	HIF1; MOP1; PASD8; HIF-1A; bHLHe78; HIF-1alpha; HIF1-ALPHA; HIF-1-alpha
NCBI Protein Information	hypoxia-inducible factor 1-alpha
UniProt Protein Name	Hypoxia-inducible factor 1-alpha
UniProt Gene Name	HIF1A
UniProt Synonym Gene Names	BHLHE78; MOP1; PASD8; HIF-1-alpha; HIF1-alpha; bHLHe78
UniProt Entry Name	HIF1A_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/hypoxia-inducible-factor-1-alpha-hif1a/188518>