

Human Interleukin 8 (IL8) ELISA Kit | IL8 elisa kit

Human Interleukin 8 (IL8) Mini Samples ELISA Kit

CATALOG # AAA188656	REACTIVITY Human	APPLICATIONS ELISA
-------------------------------	----------------------------	------------------------------


PRODUCT SYNONYMS

Interleukin 8 (IL8); Human Interleukin 8 (IL8) Mini Samples ELISA Kit; Interleukin 8; CXCL8; AMCF-I; GCP1; K60; LECT; LUCT; LYNAP; MDNCF; MONAP; NAF; NAP1; SCYB8; TSG1; B-ENAP; Neutrophil-Activating Protein 1; Granulocyte Chemotactic Protein 1; IL8 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Mini Samples Interleukin 8 (IL8). No significant cross-reactivity or interference between Mini Samples Interleukin 8 (IL8) and analogues was observed.
Sequence Length	99
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.063ng/mL
Applicable Applications for IL8 elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Mini Samples Interleukin 8 (IL8) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Mini Samples Interleukin 8 (IL8) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mini Samples Interleukin 8 (IL8). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Mini Samples Interleukin 8 (IL8). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mini Samples Interleukin 8 (IL8), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of Mini Samples Interleukin 8 (IL8) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	10834978
NCBI GeneID	3576
NCBI Accession #	NP_000575.1
NCBI GenBank Nucleotide #	NP_000575.1
UniProt Accession #	P10145
NCBI Official Full Name	interleukin-8 isoform 1
NCBI Official Synonym Full Names	C-X-C motif chemokine ligand 8
NCBI Official Symbol	CXCL8
NCBI Official Synonym Symbols	IL8; NAF; GCP1; LECT; LUCT; NAP1; GCP-1; LYNAP; MDNCF; MONAP; NAP-1; SCYB8
NCBI Protein Information	interleukin-8
UniProt Protein Name	Interleukin-8
UniProt Gene Name	IL8
UniProt Synonym Gene Names	CXCL8; IL-8; GCP-1; MDNCF; MONAP; NAP-1; LYNAP; NAF
UniProt Entry Name	IL8_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.