

General Quinolinic Acid (QA) ELISA Kit | QA elisa kit

General Quinolinic Acid (QA) Mini Samples ELISA Kit


CATALOG # AAA188761	REACTIVITY General	APPLICATIONS ELISA
-------------------------------	------------------------------	------------------------------

PRODUCT SYNONYMS

Quinolinic Acid (QA); General Quinolinic Acid (QA) Mini Samples ELISA Kit; Quinolinic Acid; QUIN; Pyridine-2; 3-Dicarboxylic Acid; QA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of Mini Samples Quinolinic Acid (QA). No significant cross-reactivity or interference between Mini Samples Quinolinic Acid (QA) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, plasma and other biological fluids
Detection Range	1.23-100ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.55ng/mL
Applicable Applications for QA elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Mini Samples Quinolinic Acid (QA) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Mini Samples Quinolinic Acid (QA) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Mini Samples Quinolinic Acid (QA) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Mini Samples Quinolinic Acid (QA) and unlabeled Mini Samples Quinolinic Acid (QA) (Standards or samples) with the pre-coated antibody specific to Mini Samples Quinolinic Acid (QA). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Mini Samples Quinolinic Acid (QA) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Mini Samples Quinolinic Acid (QA) in the sample.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/general-elisa-kits/quinolinic-acid-qa/188761>