

Human Procollagen I C-Terminal Propeptide (PICP) CLIA Kit | PICP clia kit

Human Procollagen I C-Terminal Propeptide (PICP) CLIA Kit


CATALOG # AAA188952	REACTIVITY Human	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	----------------------------	---

PRODUCT SYNONYMS

Procollagen I C-Terminal Propeptide (PICP); Human Procollagen I C-Terminal Propeptide (PICP) CLIA Kit; Procollagen I C-Terminal Propeptide; P1CP; C-Propeptide Of Type I Procollagen; Procollagen I Carboxy Terminal Propeptide; PICP clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Procollagen I C-Terminal Propeptide (PICP). No significant cross-reactivity or interference between Procollagen I C-Terminal Propeptide (PICP) and analogues was observed.
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma and other biological fluids
Detection Range	61.73-5,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 25.18pg/mL
Applicable Applications for PICP clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Procollagen I C-Terminal Propeptide (PICP) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Procollagen I C-Terminal Propeptide (PICP) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with an antibody specific to Procollagen I C-Terminal Propeptide (PICP). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Procollagen I C-Terminal Propeptide (PICP). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Procollagen I C-Terminal Propeptide (PICP) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-clia-kits/procollagen-i-c-terminal-propeptide-picp/188952>