

Human Cystatin C (Cys-C) CLIA Kit | Cys-C clia kit

Human Cystatin C (Cys-C) CLIA Kit

CATALOG # AAA188995	REACTIVITY Human	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	----------------------------	---


PRODUCT SYNONYMS

Cystatin C (Cys-C); Human Cystatin C (Cys-C) CLIA Kit; Cystatin C; CST3; Cystatin 3; Gamma Trace; Post-Gamma-Globulin; Neuroendocrine Basic Polypeptide; Amyloid Angiopathy And Cerebral Hemorrhage; Cys-C clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Cystatin C (Cys-C). No significant cross-reactivity or interference between Cystatin C (Cys-C) and analogues was observed.
Sequence Length	146
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	0.27-200ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.12ng/mL
Applicable Applications for Cys-C clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Cystatin C (Cys-C) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Cystatin C (Cys-C) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with an antibody specific to Cystatin C (Cys-C). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Cystatin C (Cys-C). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Cystatin C (Cys-C) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	118183
NCBI GeneID	1471
UniProt Accession #	P01034
NCBI Official Full Name	Cystatin-C
NCBI Official Synonym Full Names	cystatin C
NCBI Official Symbol	CST3
NCBI Official Synonym Symbols	ARMD11; HEL-S-2
NCBI Protein Information	cystatin-C
UniProt Protein Name	Cystatin-C
UniProt Gene Name	CST3
UniProt Entry Name	CYTC_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-clia-kits/cystatin-c-cys-c/188995>