

Human Myostatin (MSTN) CLIA Kit | MSTN clia kit

Human Myostatin (MSTN) CLIA Kit

CATALOG # AAA189030	REACTIVITY Human	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	----------------------------	---


PRODUCT SYNONYMS

Myostatin (MSTN); Human Myostatin (MSTN) CLIA Kit; Myostatin; GDF8; Growth Differentiation Factor 8; MSTN clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Myostatin (MSTN). No significant cross-reactivity or interference between Myostatin (MSTN) and analogues was observed.
Sequence Length	375
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	6.86-5,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 2.82pg/mL
Applicable Applications for MSTN clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Myostatin (MSTN) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Myostatin (MSTN) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with an antibody specific to Myostatin (MSTN). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Myostatin (MSTN). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Myostatin (MSTN) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	4885259
NCBI GeneID	2660
NCBI Accession #	NP_005250.1
NCBI GenBank Nucleotide #	NP_005250.1
UniProt Accession #	O14793
NCBI Official Full Name	growth/differentiation factor 8 preproprotein
NCBI Official Synonym Full Names	myostatin
NCBI Official Symbol	MSTN
NCBI Official Synonym Symbols	GDF8; MSLHP
NCBI Protein Information	growth/differentiation factor 8
UniProt Protein Name	Growth/differentiation factor 8
UniProt Gene Name	MSTN
UniProt Synonym Gene Names	GDF8; GDF-8
UniProt Entry Name	GDF8_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.