

Human Immunoglobulin Kappa (Igk) ELISA Kit | Igk elisa kit

Human Immunoglobulin Kappa (Igk) ELISA Kit

CATALOG #
AAA190136

REACTIVITY
Human

PRODUCT SYNONYMS

Immunoglobulin Kappa (Igk); Human Immunoglobulin Kappa (Igk) ELISA Kit; Igk ELISA kit (Human); Immunoglobulin Kappa; IGKC; HCAK1; Km; Immunoglobulin Kappa Constant; IgA Kappa; Igk elisa kit

PRODUCT SPECIFICATIONS


Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Immunoglobulin Kappa (Igk). No significant cross-reactivity or interference between Immunoglobulin Kappa (Igk) and analogues was observed. Reactive with Human Igk/Immunoglobulin Kappa
Assay Type	Competitive Inhibition
Samples	Serum, Plasma and other biological fluids
Detection Range	2.47-200ug/mL
Sensitivity	0.91ug/mL
Sample Size	50uL
Detection Method	Colormetric
Featured Series	CE kit
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Immunoglobulin Kappa (Igk) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Immunoglobulin Kappa (Igk) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10%, Inter-Assay: CV<12%
Function	Detects small molecule
Assay Procedure Summary	1. Prepare all reagents, samples and standards 2. Add 50ul standard or sample to each well. And then add 50ul prepared Detection Reagent A immediately. Shake and mix. Incubate 1 hour at 37 degree C 3. Aspirate and wash 3 times 4. Add 100ul prepared Detection Reagent B. Incubate 30 minutes at 37 degree C 5. Aspirate and wash 5 times 6. Add 90ul Substrate Solution. Incubate 10-20 minutes at 37 degree C 7. Add 50ul Stop Solution. Read at 450 nm immediately.
Test Principle	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Immunoglobulin Kappa (Igk) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Immunoglobulin Kappa (Igk) and unlabeled Immunoglobulin Kappa (Igk) (Standards or samples) with the pre-coated antibody specific to Immunoglobulin Kappa (Igk). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Immunoglobulin Kappa (Igk) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Immunoglobulin Kappa (Igk) in the sample.
Research Area	Immune molecule

Preparation and Storage

Store at 4 degree C for 1 month/-20 degree C for long-term (One year within shelf life)

The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/immunoglobulin-kappa-igk/190136>