

Human Free Lambda Light Chain ELISA Kit

Human Free Lambda Light Chain ELISA Kit

CATALOG #
AAA19015

REACTIVITY
Human


PRODUCT SYNONYMS

Free Lambda Light Chain; Human Free Lambda Light Chain ELISA Kit; Free Lambda Light Chain elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Assay Type	Quantitative Sandwich
Samples	Serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids
Detection Range	0.5-200mg/dl
Sensitivity	0.34mg/dl
Intra-assay Precision	Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision. Intra-Assay: CV<8%
Inter-assay Precision	Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision. CV(%) = SD/mean x 100. Inter-Assay: CV<10%
Preparation and Storage	Store at 2 to 8 degree C for 6 months.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Free Lambda Light Chain elisa kit	Principle of the Assay: This kit is an Enzyme-Linked Immunosorbent Assay (ELISA). The plate has been pre-coated with Human free L-IgLC antibody. free L-IgLC present in the sample is added and binds to antibodies coated on the wells. And then biotinylated Human free L-IgLC Antibody is added and binds to free L-IgLC in the sample. Then Streptavidin-HRP is added and binds to the Biotinylated free L-IgLC antibody. After incubation unbound Streptavidin-HRP is washed away during a washing step. Substrate solution is then added and color develops in proportion to the amount of Human free L-IgLC. The reaction is terminated by addition of acidic stop solution and absorbance is measured at 450 nm. Intended Uses: This sandwich kit is for the accurate quantitative detection of Human Free Lambda Light Chain (also known as free L-IgLC) in serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids.
--	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/free-lambda-light-chain/19015>