

## Rat Slowly Activated Delayed Rectifier Potassuim Current ELISA Kit

Rat Slowly Activated Delayed Rectifier Potassuim Current ELISA Kit

CATALOG #  
**AAA210378**

REACTIVITY  
**Rat**


### PRODUCT SYNONYMS

Slowly Activated Delayed Rectifier Potassuim Current; Rat Slowly Activated Delayed Rectifier Potassuim Current ELISA Kit; Slowly Activated Delayed Rectifier Potassuim Current elisa kit

### PRODUCT SPECIFICATIONS

|                 |                                                                   |
|-----------------|-------------------------------------------------------------------|
| Reactivity      | Rat                                                               |
| Samples         | Serum, plasma, cell culture supernate and other biological fluids |
| Assay Type      | Sandwich                                                          |
| Detection Range | 0.5ng/ml-8ng/ml                                                   |
| Sensitivity     | 0.1 ng/ml                                                         |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for Slowly Activated Delayed Rectifier Potassuim Current elisa kit

Intended Use: The Stop Solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of SLOWLY ACTIVATED DELAYED RECTIFIER POTASSUIM CURRENT in the sample, This SLOWLY ACTIVATED DELAYED RECTIFIER POTASSUIM CURRENT ELISA Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a standard curve of Optical Density versus SLOWLY ACTIVATED DELAYED RECTIFIER POTASSUIM CURRENT concentration. The concentration of SLOWLY ACTIVATED DELAYED RECTIFIER POTASSUIM CURRENT in the samples is then determined by comparing the O.D. of the samples to the standard curve.

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.