

Rabbit Osteopontin (OPN) RTU ELISA Kit | OPN rtu elisa kit

Rabbit Osteopontin (OPN) ELISA Kit

CATALOG #
AAA217570

REACTIVITY
Rabbit


PRODUCT SYNONYMS

Osteopontin (OPN); Rabbit Osteopontin (OPN) ELISA Kit; SPP1; BNSP; BSP1; ETA1; Secreted Phosphoprotein 1; Bone Sialoprotein I; Early T-Lymphocyte Activation 1; Nephropontin; Urinary stone protein; Uropontin; OPN rtu elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Specificity	This assay has high sensitivity and excellent specificity for detection of OPN. No significant cross-reactivity or interference between OPN and analogues was observed.
Sequence Length	314
Assay Type	Quantitative Sandwich
Samples	Rabbit serum, plasma, tissue homogenates or other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	< 0.065ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level OPN were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level OPN were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kits: All the reagents should be kept at 4 degree C upon receipt. For opened kits: Once the kit is opened, the remaining reagents still need to be stored according to the above storage conditions. In addition, return the unused wells to the foil pouch containing the desiccant pack and reseal along entire edge of zip-seal.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for OPN rtu elisa kit	Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to OPN. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific to OPN. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain OPN, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of OPN in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for the in vitro quantitative measurement of OPN in rabbit serum, plasma, tissue homogenates or other biological fluids.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	126723455
NCBI GeneID	100008982
NCBI Accession #	NP_001075663.1
NCBI GenBank Nucleotide #	NM_001082194.1
UniProt Accession #	P31097; P46631
Molecular Weight	35,172 Da
NCBI Official Full Name	osteopontin
NCBI Official Symbol	SPP1
NCBI Official Synonym Symbols	OPN
NCBI Protein Information	osteopontin
UniProt Protein Name	Osteopontin
UniProt Gene Name	SPP1
UniProt Synonym Gene Names	SPP-1

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.