

Rat Oxidized Low Density Lipoprotein (OxLDL) RTU ELISA Kit | OxLDL rtu elisa kit

Rat Oxidized Low Density Lipoprotein (OxLDL) ELISA Kit

CATALOG #
AAA217591

REACTIVITY
Rat


PRODUCT SYNONYMS

Oxidized Low Density Lipoprotein (OxLDL); Rat Oxidized Low Density Lipoprotein (OxLDL) ELISA Kit; OL; OxLDL rtu elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of OxLDL. No significant cross-reactivity or interference between OxLDL and analogues was observed.
Samples	Rabbit Serum, Plasma Or Other Biological Fluids
Assay Type	Quantitative Sandwich
Detection Range	3.12-200ng/mL
Sensitivity	1.33ng/mL.
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level OxLDL were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level OxLDL were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kits: All the reagents should be kept at 4 degree C upon receipt. For opened kits: Once the kit is opened, the remaining reagents still need to be stored according to the above storage conditions. In addition, return the unused wells to the foil pouch containing the desiccant pack and reseal along entire edge of zip-seal.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for OxLDL rtu elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for the in vitro quantitative measurement of OxLDL in rabbit serum, plasma or other biological fluids. Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to OxLDL. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific to OxLDL. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain OxLDL, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of OxLDL in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/rat-rtu-elisa-kits/oxidized-low-density-lipoprotein-oxldl/217591>