

Human Xanthine Oxidase ELISA Kit | XOD elisa kit

Human XOD (Xanthine Oxidase) ELISA Kit

CATALOG #
AAA22029

REACTIVITY
Human


PRODUCT SYNONYMS

Xanthine Oxidase; Human XOD (Xanthine Oxidase) ELISA Kit; XOD elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This kit recognizes Human XOD in samples. No significant cross-reactivity or interference between Human XOD and analogues was observed.
Sequence Length	513
Assay Type	Quantitative Sandwich
Samples	Serum, plasma and other biological fluids
Detection Range	0.16-10ng/mL
Sensitivity	0.1ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, mid-range and high level Human XOD were tested 20 times on one plate, respectively.
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, mid-range and high level Human XOD were tested on 3 different plates, 20 replicates in each plate.
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for XOD elisa kit	Principle of the Assay: This ELISA kit uses the Sandwich-ELISA principle. The micro-ELISA plate provided in this kit has been pre-coated with an antibody specific to Human XOD. Standards or samples are added to the micro-ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human XOD and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human XOD, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm ± 2 nm. The OD value is proportional to the concentration of Human XOD. You can calculate the concentration of Human XOD in the samples by comparing the OD of the samples to the standard curve. Intended Uses: This ELISA kit applies to the in vitro quantitative determination of Human XOD concentrations in serum, plasma and other biological fluids.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	157092381
Molecular Weight	56,177 Da
NCBI Official Full Name	xanthine oxidase
UniProt Protein Name	Xanthine oxidase
UniProt Entry Name	B3SPS5_9MICC

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/xod-human-elisa-kits/xanthine-oxidase/22029>