

General Vitamin C (VC) ELISA Kit | VC elisa kit

General Vitamin C (VC) ELISA Kit

CATALOG #
AAA22833

REACTIVITY
General


PRODUCT SYNONYMS

Vitamin C (VC); General Vitamin C (VC) ELISA Kit; AA; L-Ascorbic Acid; Ascorbate; VC elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of VC. No significant cross-reactivity or interference between VC and analogues was observed.
Assay Type	Quantitative Competitive
Samples	Serum, plasma and other biological fluids
Detection Range	493.8-40,000ng/mL
Sensitivity	< 173.5ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level VC were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level VC were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for VC elisa kit	Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to VC has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled VC and unlabeled VC (Standards or samples) with the pre-coated antibody specific to VC. After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of VC in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of VC in the sample. Intended Uses: The kit is a competitive inhibition enzyme immunoassay technique for the in vitro quantitative measurement of VC in serum, plasma and other biological fluids.
References	Comparison of some Hormonal and Biochemical Constituents in Follicular Fluid of Predominant Follicle Coexist with Oversized Follicle and its Counterpart Present in ...; IM Ghoneim

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/general-elisa-kits/vitamin-c-vc/22833>