

Human Erythropoietin (EPO) ELISA Kit | EPO elisa kit

Human Erythropoietin (EPO) ELISA Kit

CATALOG #
AAA22877

REACTIVITY
Human


PRODUCT SYNONYMS

Erythropoietin (EPO); Human Erythropoietin (EPO) ELISA Kit; EP; Epoetin; Erythropoetin; Hematopoietin; Hemopoietin; EPO elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of EPO. No significant cross-reactivity or interference between EPO and analogues was observed.
Sequence Length	193
Assay Type	Quantitative Sandwich
Samples	Human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	31.2-2,000pg/mL
Sensitivity	< 12.1pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level EPO were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level EPO were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for EPO elisa kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to EPO. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to EPO. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain EPO, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of EPO in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of EPO in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids.
---	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	62240997
NCBI GeneID	2056
NCBI Accession #	NP_000790.2
NCBI GenBank Nucleotide #	NM_000799.3
UniProt Accession #	P01588; Q2M2L6; Q549U2; Q9UDZ0; Q9UEZ5; Q9UHA0
Molecular Weight	21,307 Da
NCBI Official Full Name	erythropoietin
NCBI Official Synonym Full Names	erythropoietin
NCBI Official Symbol	EPO
NCBI Official Synonym Symbols	EP; MVCD2
NCBI Protein Information	erythropoietin
UniProt Protein Name	Erythropoietin
UniProt Gene Name	EPO

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.