

Rat Activating signal cointegor 1 complex subunit 3 (ASCC3) ELISA Kit | ASCC3 elisa kit

Rat Activating signal cointegor 1 complex subunit 3 (ASCC3) ELISA Kit

CATALOG #
AAA246365

REACTIVITY
Rat


PRODUCT SYNONYMS

Activating signal cointegor 1 complex subunit 3 (ASCC3); Rat Activating signal cointegor 1 complex subunit 3 (ASCC3) ELISA Kit; ASCC3 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	250-5000pg/mL
Sensitivity	1.0pg/ml
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



ADDITIONAL INFORMATION

Related Product Information for ASCC3 elisa kit

Intended Uses: This ASCC3 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Rat ASCC3. This ELISA kit for research use only!

Principle of the Assay: ASCC3 ELISA kit applies the competitive enzyme immunoassay technique utilizing a Polyclonal anti-ASCC3 antibody and an ASCC3-HRP conjugate. The assay sample and buffer are incubated together with ASCC3-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the ASCC3 concentration since ASCC3 from samples and ASCC3-HRP conjugate compete for the anti-ASCC3 antibody binding site. Since the number of sites is limited, as more sites are occupied by ASCC3 from the sample, fewer sites are left to bind ASCC3-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The ASCC3 concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://www.aaabiotech.com/rat-elisa-kits/activating-signal-cointegor-1-complex-subunit-3-asc3/246365>