

Mouse Oxidized High Density Lipoprotein ELISA Kit | oxHDL elisa kit

Mouse Oxidized High Density Lipoprotein ELISA Kit

CATALOG #
AAA247354

REACTIVITY
Mouse


PRODUCT SYNONYMS

Oxidized High Density Lipoprotein; Mouse Oxidized High Density Lipoprotein ELISA Kit; oxHDL elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of OXHDL. No significant cross-reactivity or interference between OXHDL and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between OXHDL and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	1.0 ng/mL
Reactivity Note	No significant cross-reactivity or interference between this antigen and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection, therefore, cross reaction may still exist in some cases.
Preparation and Storage	Store at 2-8 degree C. The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for oxHDL elisa kit	Intended Uses: This OXHDL ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Mouse OXHDL. This ELISA kit for research use only! Principle of the Assay: OXHDL ELISA kit applies the competitive enzyme immunoassay technique utilizing an anti-OXHDL antibody and an OXHDL-HRP conjugate. The assay sample and buffer are incubated together with OXHDL-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the OXHDL concentration since OXHDL from samples and OXHDL-HRP conjugate compete for the anti-OXHDL antibody binding site. Since the number of sites is limited, as more sites are occupied by OXHDL from the sample, fewer sites are left to bind OXHDL-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The OXHDL concentration in each sample is interpolated from this standard curve.
Product Categories/Family for oxHDL elisa kit	Immunology; Stem cells

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/oxhdl-mouse-elisa-kits/oxidized-high-density-lipoprotein/247354>