

Canine Hyaluronidase 2 ELISA Kit | HYAL2 elisa kit

Canine Hyaluronidase 2 ELISA Kit

CATALOG #
AAA247448

REACTIVITY
Canine


PRODUCT SYNONYMS

Hyaluronidase 2; Canine Hyaluronidase 2 ELISA Kit; HYAL2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Canine
Specificity	This assay has high sensitivity and excellent specificity for detection of HYAL2. No significant cross-reactivity or interference between HYAL2 and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between HYAL2 and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	0.1 ng/mL
Preparation and Storage	Store at 2-8 degree C. The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HYAL2 elisa kit	Intended Uses: This HYAL2 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Canine HYAL2. This ELISA kit for research use only, not for therapeutic or test applications! Principle of the Assay: HYAL2 ELISA kit applies the competitive enzyme immunoassay technique utilizing an anti-HYAL2 antibody and an HYAL2-HRP conjugate. The assay sample and buffer are incubated together with HYAL2-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the HYAL2 concentration since HYAL2 from samples and HYAL2-HRP conjugate compete for the anti-HYAL2 antibody binding site. Since the number of sites is limited, as more sites are occupied by HYAL2 from the sample, fewer sites are left to bind HYAL2-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The HYAL2 concentration in each sample is interpolated from this standard curve.
Product Categories/Family for HYAL2 elisa kit	Signal Transduction; Cancer

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/hyal2-canine-elisa-kits/hyaluronidase-2/247448>