

Mouse VLDL (Very Low Density Lipoprotein) ELISA Kit | VLDL elisa kit

Mouse VLDL (Very Low Density Lipoprotein) ELISA Kit

CATALOG #
AAA248855

REACTIVITY
Mouse


PRODUCT SYNONYMS

VLDL (Very Low Density Lipoprotein); Mouse VLDL (Very Low Density Lipoprotein) ELISA Kit; VLDL elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of VLDL. No significant cross-reactivity or interference between VLDL and analogues was observed.
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.781-50ng/ml
Sensitivity	0.469ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level VLDL were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level VLDL were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for VLDL elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-VLDL antibody was pre-coated onto 96-well plates. And the biotin conjugated anti- VLDL antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the VLDL amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of VLDL can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.