

Human GLUT2 (Glucose Transporter 2) ELISA Kit | GLUT2 elisa kit

Human GLUT2 (Glucose Transporter 2) ELISA Kit

CATALOG #
AAA271170

REACTIVITY
Human


PRODUCT SYNONYMS

GLUT2 (Glucose Transporter 2); Human GLUT2 (Glucose Transporter 2) ELISA Kit; SLC2A2; SLC2-A2; Solute carrier family 2; facilitated glucose transporter member 2; Glucose transporter type 2; liver; GLUT2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	Serum, plasma, tissue homogenates and other biological fluids.
Assay Type	Sandwich
Detection Range	1.57-100 ng/mL
Sensitivity	0.64 ng/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for GLUT2 elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Glucose Transporter 2(GLUT2). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Glucose Transporter 2(GLUT2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Glucose Transporter 2(GLUT2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Glucose Transporter 2(GLUT2) in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for GLUT2 elisa kit

Metabolic pathway; Endocrinology; Hepatology

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/glut2-glucose-transporter-2/271170>