

Mouse GAL8 (Galectin 8) ELISA Kit | GAL8 elisa kit

Mouse GAL8 (Galectin 8) ELISA Kit

CATALOG #
AAA271281

REACTIVITY
Mouse


PRODUCT SYNONYMS

GAL8 (Galectin 8); Mouse GAL8 (Galectin 8) ELISA Kit; LGALS8; PCTA1; Po66-CBP; Prostate Carcinoma Tumor Antigen 1; Lectin; Galactoside-Binding Soluble 8; Po66 carbohydrate-binding protein; Prostate carcinoma tumor antigen 1; GAL8 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Samples	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Assay Type	Sandwich
Detection Range	15.63-1000 pg/mL
Sensitivity	6.2 pg/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for GAL8 elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Galectin 8(GAL8). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Galectin 8(GAL8). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Galectin 8(GAL8), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Galectin 8(GAL8) in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for GAL8 elisa kit

Tumor immunity

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/mouse-elisa-kits/gal8-galectin-8/271281>