

Rat PON1 (Paraoxonase 1) ELISA Kit | PON1 elisa kit

Rat PON1 (Paraoxonase 1) ELISA Kit

CATALOG #
AAA271354

REACTIVITY
Rat


PRODUCT SYNONYMS

PON1 (Paraoxonase 1); Rat PON1 (Paraoxonase 1) ELISA Kit; ESA; PON; Esterase A; Serum paraoxonase/arylesterase 1; Aromatic esterase 1; A-esterase 1; Serum arylalkylphosphatase 1; PON1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Samples	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Assay Type	Sandwich
Detection Range	0.16-10 ng/mL
Sensitivity	0.066 ng/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for PON1 elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Paraoxonase 1(PON1). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Paraoxonase 1(PON1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Paraoxonase 1(PON1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Paraoxonase 1(PON1) in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for PON1 elisa kit

Metabolic pathway; Endocrinology

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/rat-elisa-kits/pon1-paraoxonase-1/271354>