

Mouse PON1 (Paraoxonase 1) ELISA Kit | PON1 elisa kit

Mouse PON1 (Paraoxonase 1) ELISA Kit

CATALOG #
AAA271728

REACTIVITY
Mouse


PRODUCT SYNONYMS

PON1 (Paraoxonase 1); Mouse PON1 (Paraoxonase 1) ELISA Kit; ESA; PON; Esterase A; Serum paraoxonase/arylesterase 1; Aromatic esterase 1; A-esterase 1; Serum arylalkylphosphatase 1; PON1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of Mouse PON1. No significant cross-reactivity or interference between Mouse PON1 and analogues was observed.
Assay Type	Sandwich
Detection Range	1.25-80 ng/mL
Sensitivity	0.51 ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay)? CV% < 8%. Three samples of known concentration were tested twenty times on 1 plate to assess intra-assay precision.
Inter-assay Precision	Inter-assay Precision (precision between assays)? CV% < 10%. Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for PON1 elisa kit

Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Paraoxonase 1(PON1). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Paraoxonase 1(PON1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Paraoxonase 1(PON1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Paraoxonase 1(PON1) in the samples is then determined by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://www.aaabiotech.com/mouse-elisa-kits/pon1-paraoxonase-1/271728>