

Mouse IDO2 (Indoleamine-2, 3-Dioxygenase 2) ELISA Kit | IDO2 elisa kit

Mouse IDO2 (Indoleamine-2, 3-Dioxygenase 2) ELISA Kit

CATALOG #
AAA274877

REACTIVITY
Mouse


PRODUCT SYNONYMS

IDO2 (Indoleamine-2; 3-Dioxygenase 2); Mouse IDO2 (Indoleamine-2; 3-Dioxygenase 2) ELISA Kit; INDOL1; Indoleamine-Pyrrole 2; 3 Dioxygenase-Like 1; Indoleamine 2; 3-dioxygenase-like protein 1; Indoleamine-pyrrole 2; IDO2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of Mouse IDO2. No significant cross-reactivity or interference between Mouse IDO2 and analogues was observed.
Assay Type	Sandwich
Detection Range	31.25-2000 pg/mL
Sensitivity	12.9 pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay)? CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.
Inter-assay Precision	Inter-assay Precision (Precision between assays)? CV%<10%. Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IDO2 elisa kit	Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Indoleamine-2,3-Dioxygenase 2(IDO2). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Indoleamine-2,3-Dioxygenase 2(IDO2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Indoleamine-2,3-Dioxygenase 2(IDO2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Indoleamine-2,3-Dioxygenase 2(IDO2) in the samples is then determined by comparing the OD of the samples to the standard curve.
Product Categories/Family for IDO2 elisa kit	Enzyme & Kinase

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/mouse-elisa-kits/ido2-indoleamine-2-3-dioxygenase-2/274877>