

Human Amyloid beta (Aggregated) ELISA Kit

Human Amyloid beta (Aggregated) ELISA Kit

CATALOG #
AAA27644

REACTIVITY
Human


PRODUCT SYNONYMS

Amyloid beta (Aggregated) ; Human Amyloid beta (Aggregated) ELISA Kit; Amyloid beta; Aggregated; Amyloid beta (Aggregated) elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	Specifically recognize Amyloid beta, no obvious cross reaction with other analogues
Assay Type	Sandwich
Samples	Serum, plasma, cell culture supernatant and other biological samples
Detection Range	78.125-5000pg/ml
Sensitivity	46.875pg/ml
Intra-assay Precision	Intra-assay Precision: samples with low, medium and high concentration are tested 20 times on same plate.
Inter-assay Precision	Inter-assay Precision: samples with low, medium and high concentration are tested 20 times on three different plates.
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Amyloid beta (Aggregated) elisa kit	Background: Amyloid beta-peptide(Abeta), a 39-43 amino acid peptide, generated by the proteolytic cleavage of amyloid precursor protein (APP) by the action of beta- and gamma-secretases. The extracellular deposition and accumulation of the generated Abeta form the main components of amyloid plaques in Alzheimer's disease. Abeta is neurotoxic, and this neurotoxicity is related to its aggregation state. Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti Amyloid beta antibody was pre-coated onto the 96-well plate. The biotin conjugated anti Amyloid beta antibody was used as the detection antibody. The standards and pilot samples were added to the wells subsequently. After incubation, unbound conjugates were removed by wash buffer. Then, biotinylated detection antibody was added to bind with Amyloid beta conjugated on coated antibody. After washing off unbound conjugates, HRP-Streptavidin was added. After a third washing, TMB substrates were added to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that turned yellow after adding a stop solution. Read the O.D. absorbance at 450nm in a microplate reader. The concentration of Amyloid beta in the sample was calculated by drawing a standard curve. The concentration of the target substance is proportional to the OD450 value.
--	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/amyloidbeta-aggregated/27644>