

Human HS (Heparan Sulfate) ELISA Kit | HS elisa kit

Human HS (Heparan Sulfate) ELISA Kit

CATALOG #
AAA277209

REACTIVITY
Human


PRODUCT SYNONYMS

HS (Heparan Sulfate); Human HS (Heparan Sulfate) ELISA Kit; HS; Heparan Sulfate; HS elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Assay Type	Competitive Inhibition
Samples	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	125-8000ng/mL
Sensitivity	51.3ng/mL
Standard	8000ng/mL
Assay Length	2h
Precision	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision. Inter-assay Precision (Precision between assays): CV%<10% Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.
Test Principle	This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Heparan Sulfate (HS) protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Heparan Sulfate (HS). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm +/- 10 nm. The concentration of Heparan Sulfate (HS) in the samples is then determined by comparing the OD of the samples to the standard curve.
Standard Curve	Matrices listed below were spiked with certain level of recombinant HS and the recovery rates were calculated by comparing the measured value to the expected amount of HS in samples. Matrix Recovery range Average serum (n=5) 84-106% 100% EDTA plasma (n=5) 87-99% 93% Heparin plasma (n=5) 87-93% 90%
Recovery Rate	The linearity of the kit was assayed by testing samples spiked with appropriate concentration of HS and their serial dilutions. The results were demonstrated by the percentage of calculated concentration to the expected. Matrix 1:21:41:81:16 serum (n=5) 79-91% 96-105% 80-92% 91-103% EDTA plasma (n=5) 92-104% 87-96% 89-101% 82-90% Heparin plasma (n=5) 84-99% 83-101% 86-99% 93-101%

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Product Categories/Family for HS elisa kit

Signal transduction; Metabolic pathway; Tumor immunity; Infection immunity; Developmental science

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/hs-heparan-sulfate/277209>