

General PGE1 (Prostaglandin E1) ELISA Kit | PGE1 elisa kit

General PGE1 (Prostaglandin E1) ELISA Kit

CATALOG #
AAA277513

REACTIVITY
General


PRODUCT SYNONYMS

PGE1 (Prostaglandin E1); General PGE1 (Prostaglandin E1) ELISA Kit; Prostaglandin E1; PGE1; PGE1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of PGE1. No significant cross-reactivity or interference between PGE1 and analogues was observed.
Assay Type	Competitive
Detection Range	31.25-2000 pg/mL
Sensitivity	9.55 pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay)? CV% < 8%. Three samples of known concentration were tested twenty times on 1 plate to assess intra-assay precision.
Inter-assay Precision	Inter-assay Precision (precision between assays)? CV% < 10%. Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for PGE1 elisa kit

Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with PGE1 protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to PGE1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of PGE1 in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for PGE1 elisa kit

Cell culture supernatant; Plasma; Saliva; Serum; Tissue Culture Media; Urine

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

