

Human HE4/WFDC2 ELISA Kit | HE4/WFDC2 elisa kit

Human HE4/WFDC2 ELISA Kit

CATALOG #
AAA282454

REACTIVITY
Human


PRODUCT SYNONYMS

HE4/WFDC2; Human HE4/WFDC2 ELISA Kit; HE4/WFDC2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay recognizes both recombinant and natural human HE4/WFDC2. The factors listed below were prepared at 16 ng/mL and assayed for cross-reactivity.
Assay Type	Quantitative Sandwich
Detection Range	125-8000pg/mL
Sensitivity	< 4.1pg/mL
Intra-assay Precision	Intra-plate Precision 3 samples with low, middle and high level HE4/WFDC2 were tested 16 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-plate Precision 3 samples with low, middle and high level HE4/WFDC2 were tested on 3 different plates, 16 replicates in each plate. Inter-Assay: CV<15%
Preparation and Storage	Ship: Ice Pack

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HE4/WFDC2 elisa kit	Background: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of HE4/WFDC2 in human serum, plasma, cell culture supernatants and other biological fluids. Principle of the Assay: This assay employs the quantitative sandwich enzyme immunoassay technique. An antibody specific for human HE4/WFDC2 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any HE4/WFDC2 present is bound by the immobilized antibody. After washing away any unbound substances, and then a detection antibody specific for HE4/WFDC2 is added to the wells and binds to the combination of capture antibody HE4/WFDC2 in sample. Following a wash to remove any unbound combination, and enzyme conjugate is added to the wells. Following incubation and wash steps, a substrate solution is added to the wells and color develops in proportion to the amount of HE4/WFDC2 bound in the initial step. The color development is stopped and the absorbance is measured.
---	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/human-elisa-kits/he4-wfdc2/282454>